



संशोधनेन संवृद्धिः

# VASANTDADA SUGAR INSTITUTE

MANJARI (Bk), 412 307, Tal. Haveli, Dist. Pune

Phone 020-26902100 (15 lines)/ 286 Fax No. : 020-26902244

Website: [www.vsisugar.com](http://www.vsisugar.com)

## COMMERCIAL BID

(Tender Notice No. 10-2018, Table Top Refrigerated Centrifuge)

TENDER FORM NO.

(Total Pages-03)

From: \_\_\_\_\_

\_\_\_\_\_

Ph/Mobile No. \_\_\_\_\_

To,  
The Director General,  
Vasantdada Sugar Institute,  
Manjari Bk., 412 307,  
Tal.- Haveli, Dist.- Pune

Sub: Tender for Table Top Refrigerated Centrifuge.

Dear Sir,

As per your tender notice published in Daily \_\_\_\_\_ dated: / /2018, we are submitting herewith our lowest offer as under.

Sr.No.	Particulars	Rate	Quantity	Amount Rs.

Note: The rate should be quoted in words & figure.

### Terms & Conditions:

1. Taxes :
2. Delivery :
3. Validity :

**Declaration:** I/We agreed to supply the instrument and complete the work within stipulated period and also accept all terms and conditions mentioned in tender.

Yours faithfully,

Signature:

Name :

Stamp of the firm:

**(Tender Notice No. 10-2018, Table Top Refrigerated Centrifuge)**

**Sub: Tender Terms & Conditions**

1. Please quote the rate including freight, all taxes, loading & unloading up to at VSI Office Manjari Bk., Haveli, Pune.
2. Demand Draft of Rs.16, 000/- of Nationalized/schedule Bank in favor of Vasantdada Sugar Institute, Pune towards Earnest Money Deposit.
3. Supplier will have to complete the order within 10 days from the date of order. If the supplier failed to complete the job within 10 days, penalty equal to 0.5 of cost of contract value per week will be recovered.
4. All Taxes, Road Tolls etc. will be on the suppliers account.
5. The Institute is not responsible for any accidents/claims during the transportation of the material.
6. Supplier should submit the delivery challan, without this unloading will not be allowed.
7. The supplier should give prior intimation before unloading.
8. Payment shall be made within 15 days after completion of order by cheque only.  
No advance is allowed.
9. The decision of Director General, VSI is binding on supplier in respect of all the dispute.
10. T.D.S. provisions of I.T. Act are applicable, and according to T.D.S. at applicable rate would be deducted from bill amount.
11. Selected agency will have to deposit the security deposit 4 % of the total order cost and sign an Agreement before placement/receiving of order on Rs.500/- non judicial stamp paper.
12. In case of down load of tender form from website, the cost of the tender form by DD should be submitted along with offer in envelope no.1 (Technical bid)
13. 95% payment will be made within 15 days against supply of material as per specification at our site. Balance 5 % amount will be made after inspection within 30 days from the date of receipt of bill.
14. If the supplier failed to supply the equipment ordered, the same would be purchased from other agency at the risk & cost of the firm.

**IMPORTANT:** The Tender should be submitted in two separate sealed envelopes in the following manner.

**Envelope No. 1. :** A D. D of EMD, Cost of Tender Form, technical manpower available, the list of customers for whom similar works have been done during last two years, a copy of registration, a copy of VAT registration, a copy of PAN card, a copy of Authorized dealership, Technical Information & brochure of the equipment offered and the list of the customers to whom similar equipments supplied.

**Envelope No. 2.:** Commercial offer only. The rates should be quoted both in words and figures. In case of variation in rate quoted in figures & word, the rate quoted in word shall be acceptable.

**Chief Accountant & I/c. Purchase**

(Tender Notice No. 10-2018, Table Top Refrigerated Centrifuge)

**Specifications**

**Qty.: 1**

**Features:**

- **Maximum speed  $\geq 30,000$  rpm**
- **Maximum RCF  $\geq 60,500xg$**
- Speed range 200 – 30,000 rpm
- **Temp range -  $-20^{\circ}C$  to  $40^{\circ}C$**
- **Dimension should be 70cm Width x 40cm Height x50cm Depth.**
- Maintenance-free induction drive
- Microprocessor with large LCD display
- Motor driven lid lock
- Automatic rotor identification system with over speed protection immediately upon rotor installation
- Active imbalance-identification and –cut off rotor specific
- Imbalance tolerant drive
- A large variety of rotors and accessories
- Audible signal at the end of each run with several melody options
- Rotor exchanger is easy to handle
- Noise level under 60 dBA at Maximum speed
- Optimized airflow system keeps the samples cool
- Simple one handed operation
- Easy to program with gloves on
- Splash proof digital display
- Permanent indication of Preset and Actual Values for rpm/rcf, indication of acceleration /deceleration, time temperature.
- Selection of speed in both rpm and g-force, with increments of 10
- 10 acceleration and 10 deceleration rates, possibility of unbroken deceleration
- Pre-selection of running time from 10 s to 99h 59min or continuous
- **Storage of up to 99 runs incl. rotor type**
- Quick-key for short runs
- Diagnostic program
- Manufacture according to international safety regulation i.e. IEC 61010
- **Angle rotor of 6 x 250 ml**
  - **Maximum speed 10,000 rpm**
  - **Maximum RCF 15000xg**
  - Adaptor for 6x 15ml
- **Angle rotor of 12 x 1.5/2.0 ml**
  - **Maximum speed 30,000rpm**
  - **Maximum RCF 60,000xg**
- Instrument should be facilitate to accommodate swing out rotor, plate rotor. Etc.

-----